Line Illuminators



Our Most Popular Configuration, Precision Engineered, Lifetime Guaranteed Film Grip

Bulbs: Four 15 watt bulbs per section.

■ Housing: Thinline 3.75" thickness.

Mounting: Surface or recess.Trim: Stainless steel.

Features: Lifetime guaranteed adjustable tension roller film grip.

Even illumination pattern. Lens removal without tools. Convection cooling system. Removable partitions.

Welded construction, white powder coated. Any length made with a continuous lens. Large, easy to use rocker switches.

Trigger / rapid start ballasts.

Optional: Low leakage ballasts.

Master switches.

Tilted lower tiers. Film activation.


Dimmers. Cords.

CARR




General

The Carr W line incorporates all the most popular configuration choices in one illuminator line. It meets all applicable standards including ACR, MIL Spec and UL. The very thin profile protrudes only 3.75" and can be recessed into the most narrow walls. To prevent inadequate light diffusion in such a thin housing, each section is equipped with no less than four 15 watt fluorescent bulbs. Featured is the patented adjustable roller film grip that carries a lifetime guarantee to hold film properly. It easily adjusts more or less firm for a smooth, positive hold. No more film dropping out because of worn or out of tolerance grips. Options include cords, film activation, master switches, dimmers, tilted lower tiers, low leakage ballasts. See Illuminator Accessories sheet for details.



Measurements, Models Numbers



Suggested Specifications

"X-ray illuminators shall be as the W Line manufactured by Carr Corporation, Santa Monica, CA (800) 952 2398. Illuminators will have four 15 watt fluorescent bulbs per section, 3 3/4" housing depth and stainless steel trim. Units shall be constructed of electro-galvanized steel with all corners welded and ground smooth, white powder coated to a high gloss finish. Film grips shall be the adjustable tension nylon roller type with lifetime guarantee. Ballasts shall be trigger / rapid start. Lens shall be easily removable for changing bulbs without the removal of any trim or screws. Illuminators shall be Underwriters Labs (UL) approved."



